

AB Tip İnceleme Sertifikası **EU Type-Examination Certificate**

Belge No / Certificate No

Belgelendirme Tarihi - Bir Sonraki Belge Tarihi /

Certification Date / Certificate Validity Date

Belge Geçerlilik Tarihi / Document Validity Period : 5 yıl / 5 years

Firma Unvanı ve Adresi /

Company Name and Address

: 92-20-03-R02

: 12.03.2021-25.12.2025

: FAGO MEDİKAL SAN. VE TİC. LTD. ŞTİ.

15 Temmuz Mah. Cami Yolu Cad. No:106 / Z1 Bağcılar/

ISTANBUL

Ürün Adı /Modeller / Product Name / Models

Direktifi / Directive

Modülü/Kategori / Module / Category

: FAGO S 101

: 2016/425 REGULATION

: B MODÜLÜ/ KATEGORİ III

MODULE B / CATEGORY III : MNA M-2020-00576, M-2021-00097, M-2021-00383

Test Rapor No/ları / Test Report No

Ürün Tipi / Product Type:

EN 149:2001+ A1:2009 Solunumla ilgili koruyucu cihazlar - Parçacıklara karşı koruma amaçlı filtreli

yarım maskeler/ Respiratory protective devices - Filtering half masks to protect against particles

Ürünün Malzeme Bilgisi / Product Material Information: FAGO S 101 model ürünleri kumaş, kulak kayışı, burun klipsi ve filtre katmanı kullanılarak imal edilmiştir./ FAGO S 101 model products are manufactured using fabric, earloop, nose clip and filter layer.

Revizyon nedeni/ Reason for revision: Farklı renkte ürünler eklenmiştir/ Different color products have been added.

Volkan AKIN 12.03.2021 Karar Verici / Approver

Okan AKEL 12.03.2021

Sirket Müdarü / General manager



MNA Laboratuvarları San. Tic.Ltd .Şti Adres: Küçükbakkalköy Mahallesi Yenidoğan Cad.No:21 Ataşehir/İstanbul Tel: 0216 574 07 08 Faks: 0216 575 13 31 www.mnalab.com




ATTACHMENTS (92-20-03-R02)

To certify the PPE product at Category III level, C2 or D module is accompanied by applying one of the conformity assessment methods along with the EU Type Examination (Module B).

Model: FAGO S 101

PPE SPECIFICATION	PERFORMANCE LEVELS
Classification	FFP2
Reusable / Single Shift Use	NR


PPE produced as a single unit to fit an individual user, all the necessary instructions for manufacturing such PPE on the basis of the approved basic model:



MNA LABORATORIES SAN. TIC. LTD. \$TÎ declares that the above-mentioned product meets the requirements of the directive according to the EU Directive 2016/425, the safety of the product is covered by the conditions and use specified in this certificate and in the technical file.



ATTACHMENTS (92-20-03-R02)





ATTACHMENTS (92-20-03-R02)



DOCUMENTS IN THE TECHNICAL FILE

- Basic Health Safety Requirements
- Risk Assessment
- Test Reports
- Technical Report

MNA Laboratuvarları San. Tic.Ltd .Şti Adres: Küçükbakkalköy Mahallesi Yenidoğan Cad.No:21 Ataşehir/ İstanbul Tel: 0216 574 07 08 Faks: 0216 575 13 31 www.mnalab.com

EU DECLARATION OF CONFORMITY

MANUFACTURER FAGO MEDİKAL SANAYİ VE TİCARET LİMİTED ŞİRKETİ

15 Temmuz Mahallesi Cami Yolu Caddesi No:106 Iç Kapı No: Z1 Bağcılar ISTANBUL / TURKEY

PRODUCT DESCRIPTION

Brand Name: Fago Model: FAGO S 101 Filtering Half Mask Class: FFP2 NR

Particle Filtering Half Face Mask in Category III product accrding to (EU) 2016/425 Personal Protective Equipment Regulation

The Manufacturer declares on his sole responsibility that the product above is, under conditions of normal use and conditions defined by the Manufacturer, safe and meets all the necessary legal conditions and requirements. The product is a personal protective equipment that is intended for single use and solely in accordance with the Manufacturer's instructions.

The Conformity is ensured with the following mechanism:

- Complies with EU 2016/425 Personel Protective Equipment Regulation establishing technical requirements for Category III products,
- Complies with Essential Health and Safety Requirements of Technical harmonised standard EN 149:2001 +A1:2009
- All required tests referred in above standards are conducted,
- Complies with other relevant harmonized legislation and community standards
- For the assessment of conformity the EU Type Examination certificate (Serial No:92-20-03) is issued, after all technical evaluations for conformity to the regulation and harmonised standards conducted, by;
 MNA LAB SAN TIC LTD STI, as Notified Body number 2841
- The product is under surveillance of same Notified Body, NB 2841 according to the Annex III (Module C2) of the PPE Regulation (EU) 2016/425, for quality assurance.

MARKING, LABELLING

Marking, labelling and user information are prepared in accordance with EU 2016/425 Personal Protective Equipment Regulation and the harmonised product standards given above.

MEASURES TO ENSURE CONFORMITY

The Manufacturer declares that he has taken all necessary measures to ensure the conformity of products placed on the market with technical documentation and technical requirements for this type of product.

GÖKHAN AYDIN

General Manager 25/12/2020

FAGO MEDIKAL SAN IVE TIC. LTD. STI.
15 Temmoz Mh. Jami Yoki Cd. Ho: 106/21
8 agcilar/157 Tig. St. No: 249684-5
Günes V. D.: 3840738071
Mersis No: U384073807100001

CE

MNA LABORATUVARLARI



TECHNICAL EVALUATION REPORT (92-20-03-R02)

Report No

: 92-20-03-R02

Report Date

: 12.03.2021

Application No

: 92-20-03

1. COMPANY INFORMATION:

FAGO MEDİKAL SAN. VE TİC. LTD. ŞTİ.

15 Temmuz Mah. Cami Yolu Cad. No:106 / Z1 Bağcılar/İSTANBUL

Tel: +90212 630 67 55 -56

E-mail: info@fagomedikal.com, birsen@fagomedikal.com


2. PPE INFORMATION:

Disposable and non-sterile half mask made of particulate protection filter material.

3. PPE TYPE IDENTIFICATION

EN 149:2001 +A1:2009 Respiratory protective devices - Filtering half masks to protect against particles - Requirements, testing, marking

4. PPE PICTURES



MNA LABORATUVARLARI



TECHNICAL EVALUATION REPORT (92-20-03-R02)



FAGO S 101

5. PPE DIMENSIONS:

FAGO S 101 model has been found to be produced using standard sizes.

6. PPE PRODUCT MATERIAL INFORMATION:

The product is made of elastic strap, nonwoven fabric on the outer and inner layers, filter material on the middle layer.

7. ESSENTIAL HEALTH AND SAFETY REQUIREMENTS

- A visual inspection was made according to EN 149:2001 +A1:2009 for ergonomics.
- Protection levels and degrees are defined by the manufacturer.
- Suitable construction materials were determined by visual inspection according to EN 149:2001 +A1:2009.
- Respiratory protective dimensions are evaluated according to EN 149:2001 +A1:2009.
- Conditioning EN 149:2001 +A1:2009 part 8.3, Penetration EN 149:2001 +A1:2009 part 8.11 (EN 13274-7), Application performance EN 149:2001 +A1:2009 part 8.4, Inward leakage EN 149:2001 +A1:2009 part 8.5, Flammability EN 149:2001 +A1:2009 part 8.6, The carbon dioxide content of the inhaled air EN 149:2001 +A1:2009 part 8.7, Inhalation resistance EN 149:2001 +A1:2009 part 8.9, Exhalation resistance EN 149:2001 +A1:2009 part 8.9 has been tested and evaluated.

MNA LABORATUVARLARI SAN. TIC. LTD. ŞTİ.

MNA LABORATUVARLARI

TECHNICAL EVALUATION REPORT (92-20-03-R02)

8. ANALYSIS AND EVALUATIONS:

EN 149:2001 +A1:2009

TESTS	PARAMETER PERFORMANCE LEVELS			RESULTS	PERFORMANCE LEVELS	EVALUATION		
		FFP1	FFP2	FFP3				
Visual inspection	Shall also the markin supplied by the manu			Appropriate	[[] []	PASS		
Banned Azo Dyes	< 30 mg/kg				< 5 mg/kg	< 30 mg/kg	PASS	
Total inward leakage	At least 46 out of the 50 individual exercise result	<25	<11	<5	See the table below	FFP2	PASS	
	At least 8 out of the 10 individual wearer arithmetic means	<22	<8	<2	See the table below	FFP2	PASS	

WHITE

rotal inwar	d Leakage (%	6)			
Exercise 1	Exercise 2	Exercise 3	Exercise 4	Exercise 5	Average
4.6	4.9	4.8	5.4	4.7	4.9
5.4	5.3	4.7	4.8	5.5	5.1
4.9	5.3	4.9	4.9		5.0
4.8	4.9	4.8			5.1
5.4	4.7	4.9			5.0
4.9	4.9	4.8			5.0
5.5	5.0	5.1	1		5.6
5.5	5.2				5.1
4.9	4.8				4.8
5.0	4.9	4.7	4.8		4.8
	Exercise 1 4.6 5.4 4.9 4.8 5.4 4.9 5.5 4.9	Exercise 1 Exercise 2 4.6	Exercise 1 Exercise 2 Exercise 3 4.6 4.9 4.8 5.4 5.3 4.7 4.9 5.3 4.9 4.8 4.9 4.8 5.4 4.7 4.9 4.9 4.8 5.5 5.0 5.1 5.5 5.2 5.5 4.9 4.8 4.9	Exercise 1 Exercise 2 Exercise 3 Exercise 4 4.6 4.9 4.8 5.4 5.4 5.3 4.7 4.8 4.9 5.3 4.9 4.9 4.8 4.9 4.8 5.4 5.4 4.7 4.9 5.0 4.9 4.9 4.8 5.4 5.5 5.0 5.1 6.0 5.5 5.2 5.5 4.7 4.9 4.8 4.9 4.6	Exercise 1 Exercise 2 Exercise 3 Exercise 4 Exercise 5 4.6 4.9 4.8 5.4 4.7 5.4 5.3 4.7 4.8 5.5 4.9 5.3 4.9 4.9 4.9 4.8 4.9 4.8 5.4 5.5 5.4 4.7 4.9 5.0 4.9 4.9 4.9 4.8 5.4 4.8 5.5 5.0 5.1 6.0 6.2 5.5 5.2 5.5 4.7 4.7 4.9 4.8 4.9 4.6 4.9

BLACK

	Total Inwar	d Leakage (%	6)			
	Exercise 1	Exercise 2	Exercise 3	Exercise 4	Exercise 5	Average
Subject 1 (As recieved)	2,6	3,5	1,7	3,0	3,1	2,8
Subject 2 (As recieved)	4,4	3,4	2,6	4,6	2,9	3,6
Subject 3 (As recieved)	4,1	1,7	2,2	2,9	2,8	2,7
Subject 4 (As recieved)	3,8	4,9	2,3	4,6	4,7	4,1
Subject 5 (As recieved)	3,7	4,4	4,2	4,7	4,8	4,4
Subject 6 (After temperature conditioning)	3,5	4,7	4,1	1,8	3,6	3,5
Subject 7 (After temperature conditioning)	3,8	4,1	2,3	2,9	5,1	3,6
Subject 8 (After temperature conditioning)	3,8	4,0	3,7	2,7	3,6	3,6
Subject 9 (After temperature conditioning)	3,9	5,1	3,5	3,6	3,8	4,0
Subject 10 (After temperature conditioning)	2,5	4,8	5,0	4,6	5,2	4,4

NA LABORATUVARLARI SAN. TÍC. LTD. STÍ.

MNA LABORATUVARLARI

TECHNICAL EVALUATION REPORT (92-20-03-R02)

TESTS PARAMETER	PARAMETER	PERFORMANCE LEVELS			RESULTS		PERFORMANCE LEVELS	EVALUATION
		FFP1	FFP2	FFP3	1			
Flammibility	Mask shall not burn burn for more than 5	n or not to continue to 5 s			Flame n	ot seen	-	PASS
Carbondioxide content of the inhalation air	Shall not exceed an a	all not exceed an average of % 1			WHITE 0,70 0,75 0,71	BLACK 0,72 0,70 0,71		PASS
Penetration of filter material	Sodium chloride, 95 % 20 % 6 % 1 See the table below %, max		e table	FFP2	PASS			
	Paraffin oil, 95 L/min %, max	% 20	% 6	%1	See the	e table	FFP2	PASS

WHITE

Penetration of filter material	Sodium Chloride (%)	Paraffin Oil (%)	
As recieved	3.6	2.8	
As recieved	3.3	3.2	
As recieved	3.5	3.0	
After the simulated wearing treatment	3.2	2.9	
After the simulated wearing treatment	3.6	2.6	
After the simulated wearing treatment	3.6	3.1	
Mechanical strength and temperature conditioning	3.4	3.1	
Mechanical strength and temperature conditioning	3.0	3.0	
Mechanical strength and temperature conditioning	3.5	3.2	

BLACK

Penetration of filter material	Sodium Chloride (%)	Paraffin Oil (%)	
As recieved	2,8	2,7	
As recieved	2,8	2,7	
As recieved	2,8	2,7	
After the simulated wearing treatment	2,7	2,8	
After the simulated wearing treatment	2,7	2,8	
After the simulated wearing treatment	2,9	2,9	
Mechanical strength and temperature conditioning	3,0	3,0	
Mechanical strength and temperature conditioning	3,0	3,0	
Mechanical strength and temperature conditioning	3,1	3,0	

NA LABORATUVARLARI SAN. TÍC, LTD. ŞTÍ.

MNA LABORATUVARLARI

TECHNICAL EVALUATION REPORT (92-20-03-R02)

TESTS PARAMETER	PARAMETER PERFORMANCE LEVELS			RESULTS	PERFORMANCE LEVELS	EVALUATION	
	FFP1	FFP2	FFP3				
Compatibility with skin		y other adverse effect		Appropriate	-	PASS	
Head harness	It can be donned and	removed easily			Appropriate	-	PASS
Breathing Resistance	Inhalation 30L/min	0,6 mbar	0,7 mbar	1 mbar	See the table below	FFP2	PASS
	Inhalation 95L/min	2,1 mbar	2,4 mbar	3 mbar	See the table below	FFP2	PASS
	Exhalation 160L/min	3 mbar	3 mbar	3 mbar	See the table below	FFP2	PASS

WHITE

Breathing Resistance (mbar)	Inhalation 30L/min (mbar)	Inhalation 95L/min (mbar)
As recieved	0.5	1.9
As recieved	0.5	1.9
As recieved	0.4	1.8
After temperature conditioning	0.4	1.8
After temperature conditioning	0.4	1.9
After temperature conditioning	0.5	1.9
After the simulated wearing treatment	0.4	1.8
After the simulated wearing treatment	0.4	1.9
After the simulated wearing treatment	0.5	1.9

BLACK

Breathing Resistance (mbar)	Inhalation 30L/min (mbar)	Inhalation 95L/min (mbar)
As recieved	0,4	1,7
As recieved	0,4	1,7
As recieved	0,4	1,8
After temperature conditioning	0,4	1,8
After temperature conditioning	0,4	1,8
After temperature conditioning	0,4	1,7
After the simulated wearing treatment	0,5	1,7
After the simulated wearing treatment	0,5	1,7
After the simulated wearing treatment	0.5	1,7

WHITE

Breathing Resistance 160L/min (mbar)	Facing directly ahead	Facing vertically upwards	Facing vertically downwards	Lying on the left side	Lying on the right side
As recieved	2,0	2,0	2,0	2,0	2,0
As recieved	1,9	2,0	2,0	2,0	2,0
As recieved	2,0	2,0	2,0	2,0	2,0
After temperature conditioning	1,9	2,0	1,9	2,0	2,0
After temperature conditioning	1,9	2,0	2,0	1,9	2,0
After temperature conditioning	1,9	2,0	2,0	2,0	2,0
After the simulated wearing treatment	1,9	1,9	1,9	2,0	2,0



MNA LABORATUVARLARI

TECHNICAL EVALUATION REPORT (92-20-03-R02)

After the simulated wearing treatment	2,0	2,0	1,9	1,9	1,9	
After the simulated wearing treatment	2,0	2,0	2,0	2,0	2,0	

BLACK

Breathing Resistance 160L/min (mbar)	Facing directly ahead	Facing vertically upwards	Facing vertically downwards	Lying on the left side	Lying on the right side
As recieved	1,8	1,8	1,8	1,8	1,9
As recieved	1,8	1,8	1,8	1,8	1,8
As recieved	1,8	1,8	1,8	1,8	1,8
After temperature conditioning	1,8	1,9	1,8	1,8	1,8
After temperature conditioning	1,8	1,9	1,8	1,8	1,8
After temperature conditioning	1,8	1,9	1,8	1,8	1,8
After the simulated wearing treatment	1,8	1,8	1,8	1,8	1,8
After the simulated wearing treatment	1,8	1,8	1,8	1,9	1,9
After the simulated wearing treatment	1,8	1,8	1,8	1,9	1,8

9. DECISION PROPOSAL

Analysis and examinations FAGO S 101 model coded personal protective equipment; Respiratory Protective Devices EN 149:2001 +A1:2009- Filtered Half Masks for Protection Against Particles - Properties, Experiments and Marking standards are evaluated. It is recommended to be certified at the performance levels specified as a result of technical evaluations.

10. ATTACHMENTS

- Basic Health Safety Requirements
- Risk Assessment
- Test Reports
- User Instruction

Reason for revision : Different color products have been added.

CONTROLLER : VOLKAN AKIN

SING :

DATE : 12.03.2021